

UNLV RMS

UNLV Guidelin

UNLV Guideline for Risk Assessment and Control of Radiation Producing Devices

Medical
1 An individual working outside the device while the beam is on could exceed the maximum allowable dose to a member of the public. No exposure from working inside of the device.

Reason: The beam is not enclosed. However, no work allowed inside the device while the beam is on.

The beam path is not enclosed.

Shielding and external surfaces that are fixed in place or are interlocked to control X-ray production

Radiation levels are may exceed 50 microR at in 1 hour at 30 cm from the surface of the device under all operating conditions

Devices are certified by US FDA for use with human patients

The device must be:

1. registered annually with the State of NV
2. authorized by Radiation Safety prior to operation
3. surveyed by Radiation Safety prior to initial operation and re-surveyed prior